

MAR 2 8 2002

## 

### **General Information**

**Submitters Name/Address:** 

ArthroCare, Corporation

680 Vaqueros Ave

Sunnyvale, CA 94085-2936

**Establishment Registration Number:** 

2951580

**Contact Person:** 

Valerie Defiesta-Ng

Director, Regulatory Affairs

Phone Number:

(408) 736-0224

Date Prepared:

February 25, 2002

**Device Description** 

Trade Name:

ArthroCare® Electrosurgery Wands

Generic/Common Name:

Electrosurgical Device and Accessories

**Classification Name:** 

Electrosurgical Cutting and Coagulation

Device and Accessories (21 CFR

878.4400)

#### **Predicate Devices**

ArthroCare Electrosurgery System Wands

K014222; cleared on January 18, 2002

#### **Product Description**

The Wands are bipolar, single use, high frequency electrosurgical devices designed for specific indications.

### **Intended Use**

The Electrosurgery Wands are indicated for resection, ablation, and coagulation of soft tissue and hemostasis of blood vessels in open, laparoscopic, and endoscopic general surgery and general gynecology procedures. Representative procedures include the following:

General Surgery 1 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2
cholecystectomy
lysis of adhesions
upper GI
GI (other)
splenectomy
thyroidectomy
herniorrhaphy
breast biopsy
bowel resection
pelvic adhesiolysis
removal of lesions
removal of polyps
tumor biopsy
Gynecological Surgery A Part 1888
lysis of adhesions
hysterectomy
salpingo-oophorectomy
burch colposuspension
myomectomy
endometriosis
ovariohysterectomy
removal of tumors

### Substantial Equivalence

This Special 510(k) proposes a modification in dimensional and performance specifications, materials, and labeling for the Wands, which were previously cleared under K014222 on January 18, 2002. The indications for use, technology, principle of operation, packaging, and sterilization parameters of the Wands remain the same as in the predicate cleared 510(k)s.

### **Summary of Safety and Effectiveness**

The Electrosurgery System modified Wands, as described in this submission, are substantially equivalent to the predicate, unmodified Wands. The proposed modification in dimensional and performance specifications, materials, and labeling are not substantial changes or modifications, and do not significantly affect the safety or efficacy of the devices.



Food and Drug Administration 9200 Corporate Boulevard Rockville MD 20850

# MAR 2 8 2002

ArthoCare Corporation
Ms. Valerie Defiesta-Ng
Director, Regulatory Affairs
680 Vaqueros Avenue
Sunnyvale, California 94085-2936

Re: K020622

Trade Name: ArthroCare® Electrosurgery Wands

Regulation Number: 878.4400

Regulation Name: Electrosurgical cutting and coagulation device and accessories;

Regulatory Class: II Product Code: GEI

Dated: February 26, 2002 Received: February 26, 2002

### Dear Ms. Defiesta-Ng:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

### Page 2 - Ms. Valerie Defiesta-Ng

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4659. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address http://www.fda.gov/cdrh/dsma/dsmamain.html

Sincerely yours,

Mulain C. Provost Cor Celia M. Witten, Ph.D., M.D.

Director

Division of General, Restorative and Neurological Devices Office of Device Evaluation Center for Devices and Radiological Health

Enclosure

### **Indications for Use Statement**

Device Name:

ArthroCare® Electrosurgery Wands

510(k) Number:

К<u>ОДО6Д</u>Д

Indications for use:

The Electrosurgery Wands are indicated for resection, ablation, and coagulation of soft tissue and hemostasis of blood vessels in open, laparoscopic, and endoscopic general surgery and general gynecology procedures. Representative procedures include the

following:

General Surgery
cholecystectomy
lysis of adhesions
upper GI
GI (other)
splenectomy
thyroidectomy
herniorrhaphy
breast biopsy
bowel resection
pelvic adhesiolysis
removal of lesions
removal of polyps
tumor biopsy
Gynecological Surgery
lysis of adhesions
hysterectomy
salpingo-oophorectomy
burch colposuspension
myomectomy
endometriosis
ovariohysterectomy
removal of tumors

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription U	se
(Per 21 CFR 8	01.109)

\_\_ OR Over-the-Counter Use

(Division Sign-Off)

Division of General, Restorative and Neurological Devices

510(k) Number <u>K020622</u>